

## National Voluntary Laboratory Accreditation Program



### SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

**Tovey Engineering, Inc.** 

22602 N. 17th Avenue Phoenix, AZ 85027-1303 Mr. Michael Tovey

Phone: 623-434-5110 Fax: 623-434-5130 E-mail: miketovey@toveyengineering.com URL: www.toveyengineering.com

## **CALIBRATION LABORATORIES**

#### **NVLAP LAB CODE 200662-0**

NVLAP Code: 20/A01 ANSI/NCSL Z540-1-1994; Part 1 Compliant

**MECHANICAL** 

NVLAP Code: 20/M06

Force

Free Weights

Range Best Uncertainty  $(\pm)$  in ppm note 1 Remarks

1 gf to 50 lbf Tension & Compression

Dead Weight Method

Range in lbf Best Uncertainty  $(\pm)$  in ppm note 1 Remarks

1 to 1000 Tension & Compression

Transfer Standard Method

Range in lbf Best Uncertainty (±) in % note 1 Remarks

 $\begin{array}{cccc} 10 \text{ to } 110 \text{ k} & 0.025 & \text{Tension \& Compression} \\ 50 \text{ k to } 1000 \text{ k} & 0.05 & \text{Compression} \\ 50 \text{ k to } 800 \text{ k} & 0.05 & \text{Tension} \end{array}$ 

Indicators - DC mV/V voltage ratio measurement

Range in mV/V Best Uncertainty  $(\pm)$  in  $\%^{note 1}$  Remarks

0 to 10 0.003

2006-04-01 through 2007-03-31

Effective dates

For the National Institute of Standards and Technology

NVLAP-01S (REV. 2004-10-31)

Page 1 of 2



# National Voluntary Laboratory Accreditation Program



## **CALIBRATION LABORATORIES**

**NVLAP LAB CODE 200662-0** 

2006-04-01 through 2007-03-31

Effective dates

CN Laisin

For the National Institute of Standards and Technology

<sup>1.</sup> Represents an expanded uncertainty using a coverage factor, k=2, at an approximate level of confidence of 95 %.

<sup>2.</sup> Typical uncertainties may be up to 0.05 %.